

<b><u>Ordination Equipment</u></b> <b>according to MED.D.010 (c) (1) Regulation (EU) No. 1178/2011</b> <b>and ATCO.MED.C.010 (c) (1) Regulation (EU) No. 2015/340</b>				
<b>Clinical examination:</b>	<input checked="" type="checkbox"/> Stethoscope	<input type="checkbox"/> BP Monitore	<input type="checkbox"/> calibrated Scale	<input type="checkbox"/> Tapeline
<b>Cardiological examination:</b>	<input checked="" type="checkbox"/> 12-lead-ECG	<input type="checkbox"/> Spirometry <sup>4</sup>		
<b>Urin analysis:<sup>1</sup></b>	<input type="checkbox"/> Urine test			
<b>Laboratory:<sup>2</sup></b>	<input type="checkbox"/> Serum Lipids	<input type="checkbox"/> Haemoglobin		
<b>Eye-Examination:</b>	<input type="checkbox"/> determination of near, intermediate and distant vision <sup>3</sup>		<input type="checkbox"/> Ishihara plates <sup>3</sup> (24 plates-Version)	<input type="checkbox"/> Ophthalmoscope
<b>ENT-examination:</b>	<input checked="" type="checkbox"/> pure-tone-audiometry	<input type="checkbox"/> Otoscope	<input type="checkbox"/> Rhinoscope	

<sup>1</sup> minimum scope: GLU, ERY, PROT

<sup>2</sup> blood sampling equipment obligatory, as well as cooperation with a laboratory

<sup>3</sup> requirements VO(E)U no 1178/2011 MED.B.070 according GM MED.B.070; possibly, in the case of artificial lighting, daylight spectrum should be provided (C or D65)

<sup>4</sup> if not available, cooperation with a specialist in pulmonology